

High Performance Liquid Chromatography

Application News

No.**L510**

Analysis of Residual Antimicrobials in Meat with Antimicrobial Screening System (Part 2)

Antimicrobials are a type of veterinary drug and animal feed additive, and are used for the treatment and prevention of disease in livestock and marine products. Residual antimicrobials are often found in livestock and marine products, so threshold levels for antimicrobials are set by regulation to ensure the safety of the consumer based on amounts that do not harm human health.

Due to ongoing reports of recent cases of regulatory violations in various countries and the large number of compounds targeted for testing, there is a demand for quick and simple antimicrobial screening.

While Application News No.L509 described an example of using the antimicrobial screening system for screening 12 quinolone compounds, this Application News describes an example screening analysis of 12 antimicrobial target compounds including sulfanomides.

Sample Pretreatment

Sample pretreatment for analysis of residual antimicrobials in meat usually employs liquid-liquid extraction (and sometimes solid phase extraction), but this process takes time and effort. In this article, we employed a QuEChERS method designed to be more efficient and reduce pretreatment times. The QuEChERS method is used to pretreat vegetables and fruits for residual pesticide analysis.

After using the QuEChERS method to perform extraction and fat removal, sample solutions were prepared by evaporation and redissolution steps. Table 1 shows the maximum residue limits (MRLs) of target compounds and sample solution concentrations after sample pretreatment, and Fig. 1 shows the sample pretreatment protocol. Refer to the instruction manual of the system for the details of the sample pretreatment procedure.

Table 1	able 1 Maximum Residue Limits and Sample Solution		
	Concentration of Screening Target Compounds		

	Compound	MRL	Sample Solution
	Compound	(mg/kg)	Concentration (mg/L)
1	Sulfadiazine	0.01	0.025
2	Sulfamerazine	0.01	0.025
3	Sulfadimidine	0.01	0.025
4	Sulfamonomethoxine	0.01	0.025
5	Trimethoprim	0.01	0.025
6	Sulfamethoxazole	0.01	0.025
7	Ormetoprim	0.01	0.025
8	Sulfadimethoxine	0.01	0.025
9	Sulfaquinoxaline	0.01	0.025
10	Pyrimethamine	0.01	0.025
11	Difurazon	0.01	0.025
12	Nicarbazin*1	0.01	0.025

*1: Concentration of N, N'-Bis(4-nitrophenyl)urea, the main constituent of nicarbazin.




Fig. 1 Sample Pretreatment Protocol

Analysis of Antimicrobials Including Sulfonamides in Meat

Chicken and beef were used as samples. The analytical conditions are shown in Table 2. Chromatograms of the pretreated matrix solutions (blue line), matrix solutions spiked with standard solution to create matrix standard solutions (red line), and neat standard solution (black line) are shown in Fig. 2.

Standard solution was added to matrix solutions to make up antimicrobial concentrations, including sulfonamide concentrations, of 0.01 mg/kg in matrix standard solutions. Standard solutions were prepared to the sample solution concentrations listed in Table 1.


The photodiode array (PDA) detector (six-wavelength) built in the i-Series instrument was used for detecting all target compounds. Employing the analytical conditions shown, all 12 compounds were separated and eluted in approximately 25 minutes.

Table 2 Analytical Conditions


: LC-2040C 3D
: Shim-pack FC-ODS (150 mm L. × 4.6 mm I.D., 3 μm)
: A) 20 mM (Sodium) Phosphate Buffer Containing 0.1 M Sodium Perchlorate B) Acetonitrile/Methanol=80/20
,
: Gradient Elution
: 1.0 mL/min
: 50 °C
e : 20 μL
: 240 nm
270 nm
280 nm
285 nm
350 nm
380 nm
: 40 °C


Similarity Calculation Using UV Spectral Library

All target compounds in this Application News can be analyzed qualitatively based on UV spectra as well as retention times. Sample spectra can be also checked for similarity against library spectra. Fig. 3 shows a UV spectrum of sulfaquinoxaline in a beef matrix spiked with a standard solution of sulfaquinoxaline at threshold concentration. Degree of similarity with the library spectrum was 0.997.









Matrix Standard Solution (Red Line), Matrix Solution (Blue Line), Neat Standard Solution (Black Line)

First Edition: Oct. 2016



Shimadzu Corporation www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.

This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.

The content of this publication shall not be reproduced, altered or sold for any commercial purpose without the written approval of Shimadzu. Company names, product/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation or its affiliates, whether or not they are used with trademark symbol "TM" or "®". Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services. Shimadzu disclaims any proprietary interest in trademarks and trade names of ther than its own.

The information contained herein is provided to you "as is" without warranty of any kind including without limitation warranties as to its accuracy or completeness. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication. This publication is based upon the information available to Shimadzu on or before the date of publication, and subject to change without notice.